



State of the Science Review

The use of smart environments and robots for infection prevention control: A systematic literature review



Davide Piaggio PhD^{a,*}, Marianna Zarro^{a,b}, Silvio Pagliara PhD^a, Martina Andellini PhD^a, Abdulaziz Almuhini MSc^{a,e}, Alessia Maccaro PhD^a, Leandro Pecchia PhD^{a,c,d}

^a School of Engineering, University of Warwick, Coventry, UK

^b Department of Internal Medicine and Medical Therapy, University of Pavia, Pavia, Italy

^c Università Campus Bio-Medico, Roma, Italy

^d R&D Blueprint and COVID-19, World Health Organization, Genève, Switzerland

^e Department of Biomedical Technology, College of Applied Medical Sciences, Prince Sattam Bin Abdulaziz University, Al-Kharj, Saudi Arabia

Key Words:

Robot
Artificial intelligence
Internet of things
Hand hygiene
Health 4.0

A B S T R A C T

Background: Infection prevention and control (IPC) is essential to prevent nosocomial infections. This manuscript aims at investigating the current use and role of robots and smart environments on IPC systems in nosocomial settings

Methods: The systematic literature review was performed following the PRISMA statement. Literature was searched for articles published in the period January 2016 to October 2022. Two authors determined the eligibility of the papers, with conflicting decisions being mitigated by a third. Relevant data was then extracted using an ad-hoc extraction table to facilitate the analysis and narrative synthesis.

Results: The search strategy returned 1520 citations and 17 papers were included. This review identified 3 main areas of interest: hand hygiene and personal protective equipment compliance, automatic infection cluster detection and environments cleaning (ie, air quality control, sterilization). This review demonstrates that IPC practices within hospitals mostly do not rely on automation and robotic technology, and few advancements have been made in this field.

Conclusions: Increasing the awareness of healthcare workers on these technologies, through training and involving them in the design process, is essential to accomplish the Health 4.0 transformation. Research priorities should also be considering how to implement similar or more contextualized alternatives for low-income countries.

© 2023 The Author(s). Published by Elsevier Inc. on behalf of Association for Professionals in Infection Control and Epidemiology, Inc. This is an open access article under the CC BY license (<http://creativecommons.org/licenses/by/4.0/>)

At the end of 2019, anomalous pneumonia cases were identified in Wuhan City (China) and were reported to the World Health Organization (WHO) on 31 December 2019. Scientists were able to identify the cause of these atypical pneumonia cases: a new strain of coronavirus, later renamed as SARS-CoV-2. From then on, the situation rapidly degenerated until when, a few months later, the WHO declared the disease caused by SARS-CoV-2 (COVID-19) a global pandemic. As of 25 October 2022, there have been 625 million confirmed

cases of COVID-19 globally, and 6.5 million deaths have been reported by the WHO.

The geographic distribution of COVID-19 cases is uneven, with Europe still being the most affected WHO Region (totaling 260 million of confirmed cases, against the 9 million confirmed cases in the Africa WHO Region).¹ Multiple attempts to explain these differences have been made, that is, genetic immunity, climatic conditions, population' age, different screening activities,² as well as undertesting and underreporting.^{3,4} One thing is certain, higher resource settings found themselves, for the first time after World War II, in conditions of deprived resources, typical of lower resource settings (eg, scarcity of ventilators, personal protective equipment (PPE), beds in hospitals, health care workers, etc).⁵

* Address correspondence to Davide Piaggio, PhD, School of Engineering, University of Warwick, Library Road, CV47AL, Coventry, UK.

E-mail address: davide.piaggio@warwick.ac.uk (D. Piaggio).

Conflict of interest: The authors declare that they have no competing interests.

Availability of data and materials: The datasets used and/or analyzed during this study are available from the corresponding author on reasonable request.

One of the key lessons learned from COVID-19 was the extreme urgency to better prepare for future pandemics, with prevention and containment measures enlisted as top priorities of both political leaders and scientists. Certain circumstances, such as the climate crisis and the growing global population, will, in fact, likely lead to new pandemics, due to the increased risk of zoonoses. These reasons reinforce the interest of the scientific community in developing new and more efficient solutions to such threats. This paper will help showcase new approaches developed and tested for infection prevention and control (IPC) purposes in particular in the field of robotics and automation, and hopefully will help improve hospital responses and preparedness to forthcoming outbreaks.

Various studies^{6,7} highlighted the weaknesses of IPC guidelines, which did not ensure an effective response to the COVID-19 pandemic. In 2022, in hindsight, the Lancet Commission highlighted the key components needed for effective preparedness plans, one of which is the prompt adoption of IPC procedures designed following the most updated knowledge on transmission routes of respiratory infectious diseases.⁷ On the other hand, the WHO published a statement in response to the Lancet Commission, chronologically showing that their approach during the first waves of COVID-19 was as adequate and efficient as possible.⁸ Therefore, prompt innovation of IPC strategies remains crucial to improve preparedness and health care response to pandemics. In fact, IPC was also pinpointed as a research priority by the Global Research Forum, organized by the WHO, at the start of the pandemic, in February 2020. After almost 2 years' worth of work, which saw the joint effort and collaboration of international experts with different backgrounds, in December 2021, the WHO published the document "Infection prevention and control in the context of coronavirus disease (COVID-19): A living guideline", replacing an old version dating back to December 2020. The most recent version was last updated on April 2022, mostly revising the advice on mask use for children, thanks to the United Nations Children's Fund (UNICEF) contribution.⁹

Technological innovation is playing a leading role in the "new approach" of Health 4.0 and certainly offers promising solutions for hospital responses to future epidemic/pandemic outbreaks. The shift in health care driven by Health 4.0 is based on the integration of Internet of Things, Cloud and Fog Computing, and Big Data.¹⁰ It is clear that automation, digital tools and robots could have played a key role in the IPC in health care settings, and they should therefore be leveraged to ensure more crucial roles in this remit. ODIN, one of the biggest trailblazing Horizon 2020 projects in the field of robotics applied to health care, is already looking into this direction, aiming to enhance hospital safety, productivity and quality relying on artificial-intelligence-based technologies.¹¹

The purpose of this paper was to conduct a systematic literature review on the use of robots and automation for IPC purposes in health care settings, focusing on nosocomial infections. We intended to investigate the performance measures, the health care workers (HCWs) compliance, as well as cost and resources needed (including personnel, time, infrastructure, existing servers/computer systems, etc) of IPC technologies compared with the gold standard of practice, if existent. This is of vital importance for informing the preparedness plans to tackle the next global health emergency.

METHODS

The methodology that was applied followed the PRISMA statement for systematic literature reviews.¹²

Search strategy

In order to carry out the systematic literature review, significant keywords were selected and included in a search string. Such

keywords were identified by reading relevant infection prevention and control (IPC) literature, in particular the World Health Organization (WHO) Guidelines for Infection prevention and control guidance for long-term care facilities in the context of COVID-19.¹³ These were then reviewed and discussed by the Applied Biomedical Signal Processing and Intelligent eHealth (ABSPIeH) lab members, composed of biomedical engineers, computer scientists, biologists, and bioethics experts. These terms, which were combined using Boolean operators (eg, AND, OR), are hereby reported: "Infection prevention and control", "Infection prevention and control program", "infection prevention", "infection control", "transmission control", "health care", "assistive care", "hospital", "robot", "telerobotic", "teleoperation", "automated technology", "artificial intelligence", "machine learning", "deep learning", "internet of things", "smart device", "smart service", "smart technology", "sensor", "wearable", "key enabling technology", "human support robot". The full search string can be found in [Supplementary Figure A1](#). This search was performed on OVIDSp for the period January 2016 to October 2022. This was originally searched in 2021 limiting the search to the previous 5 years. The search was then expanded in a second stage by looking for all the contributions between March 2021 and October 2022.

Inclusion/exclusion criteria

We judged eligible peer reviewed journal articles, written in English, that reported on automation technologies for Infection Prevention and Control (IPC) in health care settings. Letters to editors, book chapters, editorials and notes were excluded. Reviews were also excluded, and they were only screened to check if there were any recent ones analogous to our review. No recent review dealt with the same topic. Moreover, articles that dealt with patients' screening, diagnosis and other procedures not relevant to IPC were excluded. The articles were appraised by 2 authors by title, abstract, and, finally, full text, while a third 1 independently reviewed the results of the screening. Disagreements among the 2 authors were solved by arbitration of the third.

Data extraction and quality appraisal

Relevant data was extracted and collected in an ad-hoc Excel sheet (see [Supplementary Table A1](#)). The extracted data was organized by 5 macro areas, that is, broad thematic areas: (1) hand hygiene compliance, (2) cleaning and disinfection of hospital environment, (3) infection cluster detection, (4) air quality control, (5) correct use of PPE. The quality appraisal was conducted using the MMAT tool,¹⁴ due to the mixed types of studies considered. The MMAT tool contains different subsections that allow for the assessment of both quantitative and qualitative studies.

Data synthesis

To synthesize the data extracted and collected we used the narrative synthesis method.^{15,16} For each paper we identified the focus of interest, which was then described in the result section and organized based on principal themes (i.e., interpretive approach). This information is then used to formulate discussions and possible solutions.¹⁶

RESULTS

Search outcome

The OvidSP search and study selection process is summarized in [Figure 1](#).




Fig 1. PRISMA flow diagram. Study selection process used, divided into 3 phases: identification, screening, included.

The aforementioned combined searches returned 1,520 hits, 21 of which were duplicates from the previous search; thus, they were excluded. Overall, 17 articles met our inclusion criteria (Supplementary Table A2 reports the reasons for exclusions for the full text screening round).

Data extraction

Table 1 summarizes the essential information, including the IPC device used, the aim of the study, macro area, the participants for each study and hospital department.

Out of the 17 included studies, 11 focused on hand hygiene (HH) compliance of HCWs,^{17–27} 2 reported on automatic cluster detection systems.^{28,29} Four articles addressed different streams, namely air quality control,³⁰ personal protective equipment (PPE) compliance,³¹ and hospital spaces' cleaning and disinfection.^{32,33} An overview of the technologies presented in the included studies is shown in Figure 2.

The rest of this section will summarize and group the results by the aforementioned 5 macro areas.

Hand hygiene compliance

As previously stated, most of the studies investigated the use of automated technologies for monitoring and encouraging hand hygiene compliance of HCWs. Eight of these studies tackled the problem using wearable sensors, with some of them including signaling

devices that reminded HCWs to complete the task of sanitizing their hands.^{17–19,22–25,27}

McCalla et al¹⁸ developed an automated hand hygiene compliance system (HHCS), based on sound and light signals activated when a HH opportunity is detected. In their first article (2017) they assessed the effects of HHCS on HCWs' hand hygiene in the intensive care unit, comparing it to the gold standard, that is, human observation. They found that, despite a higher number of HH opportunities recorded, there was a lower number of actual HH events. In 2018, they implemented and pilot tested the same system in the whole hospital.¹⁹ In this case, they observed a significant reduction in the rate of catheter-associated urinary tract infections and of central line-associated bloodstream infections. In both studies the Biovigil system (Biovigil Healthcare Systems, Inc) was used. This system is the only electronic hand hygiene monitoring solution that provides a reminder to sanitize hands, reassuring patients and anybody else at the bedside that hand hygiene has been performed.

Similarly, Xu et al used an Internet of Things HH monitoring device during the study period. To monitor the process, they installed 5 transmitters connected to an IoT gateway wirelessly. In contrast with McCalla et al, they observed a drastic increase of HH compliance rate. However, the infection rates of the hospital were not significantly different.¹⁷ Xu Q et al adopted the Sanibit system to assess HH compliance and behaviors among HCWs in the surgical intensive care unit. The Sanibit system works thanks to room sensors and Bluetooth wristbands, which detect HH opportunities and monitor HH compliance and quality. The system is enhanced by real-time feedback via the wristbands, a gamification app where each HCW can check their performance, and an automated HH compliance analysis.

Table 1
Study characteristics

Study	Macro area	IPC device	One-sentence aim of the study	Participants	Hospital department/area
Xu 2021 ¹⁷	Hand hygiene compliance	IOT hand hygiene compliance monitoring device	Evaluation of IPC device impact on hand hygiene (HH) compliance and health care-associated infection rates	Hospital staff (54): specialized doctors, doctors, nurses, and cleaners Patients (697)	Electronic Intensive Care Unit (EICU)
McCalla 2017 ¹⁸	Hand hygiene compliance	Hand hygiene compliance system - Biovigil Healthcare Systems Inc	Evaluation of IPC device impact on HH compliance	Hospital staff: nurses, nurse technicians, respiratory therapist, care managers, dietary aides, housekeeping staff Patients (4,070)	Intensive Care Unit (ICU)
McCalla 2018 ¹⁹	Hand hygiene compliance	Hand hygiene compliance system - Biovigil Healthcare Systems Inc	Evaluation of IPC device impact on health care-associated infection rates	Hospital staff: nurses, nurse technicians, respiratory therapist, care managers, dietary aides, housekeeping staff Patients (36,890)	Whole hospital
Edmisten 2017 ²⁰	Hand hygiene compliance	Electronic HH monitoring system, based on radiofrequency	Report on IPC device implementation, challenges, and success	Hospital staff (2,830)	Three community hospitals
Dufour 2017 ²¹	Hand hygiene compliance	Electronic HH monitoring system, based on radiofrequency	Report on HH compliance	Hospital staff (42): 23 medical doctors, 8 residents, 12 medical students, 3 senior doctors, 6 nurses, 9 assistant nurses and 4 housekeepers	Seven patient rooms, unit not specified
Iversen 2020 ²²	Hand hygiene compliance	HHC automated monitoring system (Sani nudge)	Evaluation of HH compliance	Hospital staff: 42 nurses	Orthopedic surgery department, oncology department
Xu 2021 ²³	Hand hygiene compliance	Electronic HH system - Sanibit	Validation of IPC device	Hospital staff (15): 12 nurses, 2 patient care assistants and 1 secretary	Surgical intensive care unit
Xu 2022 ²⁴	Hand hygiene compliance	Electronic HH system - Sanibit	Evaluation of HH individual behaviors	Hospital staff (15): 12 nurses, 2 patient care assistants and 1 secretary	Surgical intensive care unit
Akkoc 2021 ²⁵	Hand hygiene compliance	Electronic hand hygiene reminding and recording systems (EHHRRSs)	Validation of IPC device	Hospital staff: nurses, physicians, transporters, and other staff Patients (248)	Anesthesia and reanimation intensive care unit
Huang 2021 ²⁶	Hand hygiene compliance	automatic hand hygiene monitoring system (MediHandTrace), based on radiofrequency	Evaluation of IPC device impact on HH compliance	Hospital staff: 38 physicians, 13 interns, 37 nurses, 18 nursing assistants, and 5 housekeeping personnel	Infection unit
Durant 2020 ²⁷	Hand hygiene compliance	Electronic hand hygiene monitoring systems (EHHMS)	Report on New York State hospitals' adoption of EHHMS. Evaluation of IPC device on C. Difficile infection rates	56 hospitals	Not relevant
Stachel 2017 ²⁸	Infection cluster detection	Statistical software SaTScan and software for laboratory data management WHONET	Report on IPC device implementation	Patients	Two hospitals
Aghdassi 2021 ²⁹	Infection cluster detection	automated cluster alert system (CLAR)	Report on IPC device implementation and on cluster detected	Patients	Whole hospital
Colella 2022 ³⁰	Air quality control	Operating room air quality monitoring system based on fuzzy logic (FL)	Report on IPC device development	Hospital staff, Patients	Operating room (OR)
Preda 2022 ³¹	Correct use of PPE	Artificial intelligence- personal protective equipment (AI-PPE) compliance system	Validation of IPC device	Hospital staff (74): 6 nurses, 14 medical students, 3 physicians, 9 junior medical officer, 3 surgeons, 31 laboratory staff and 8 administrative staff	Not specified
Wang 2022 ³²	Cleaning and disinfection of hospital environments	RNN neural networks with the addition of PDCA cycle related element	Evaluation of IPC device impact on workers' satisfaction and standardization rates	Hospital staff: 17 room nurses	Supply room
Khan 2020 ³³	Cleaning and disinfection of hospital environments	Different types of robotic technologies are used in hospital setting to dry vacuum and mopping to remove germs and pesticides. - intelligent navigating vacuum pump - ultra-violet radiation based device - highly dynamic robotic gripper and sensing system - autonomous heavy-duty cleaning robot	Report on robot utilization to manage the COVID-19 pandemic	Not relevant	Not relevant




Fig 2. Technologies employed by the studies included: wearable sensor with and without reminders (7), artificial intelligence (AI) (fuzzy logic, neural networks) (3), radio-frequency identification technology (3), automated cluster alert system (2), Internet of Things (1), robot (1).

In 2021, they published the results of the validation of the technology and they also reported a significant difference between individual HCWs' HH behaviors.²³ In 2022, they published the results of an expansion of these observations using the same systems and participants. In this case, they report higher compliance rates when exiting the patient room and after long visits compared to the compliance rate when entering the patient room or after short visits.²⁴ The differences in individual HH behavior patterns suggest that personalized interventions could improve HH compliance.

Similar to the work of McCalla et al and Akkoc et al found a significant reduction of health care-related infections, specifically central line-associated bloodstream infections and a ventilator-associated pneumonia, when considering the effects of tools to raise HH awareness. In particular, they analyzed the effect of an electronic HH recording and reminder system (EHHRRSs) on nosocomial infection rates, comparing it to the conventional observation method.²⁵ The authors also point out that the discontent of HCWs regarding the use of the tracking device limited the duration of the study.

Regarding the adoption of such tools, Durant et al found, through surveys and interviews, that the number of hospitals adopting an Electronic HH Monitoring System (EHHMS) in the New York State area was low, mostly due to cost and concerns on the accuracy of the devices.²⁷ Moreover, their analysis on the EHHMS' impact on hospital-acquired *Clostridium difficile* shows no significant effect. The sanitizers' location was proved to be key to HH compliance by Iversen et al, using an automated monitoring system to evaluate HH compliance. They also could not find any association between HH compliance and the number of beds in rooms.²²

A different approach was adopted by 3 of the 11 articles dealing with HH compliance, using devices based on radio-frequency identification (RFID) technologies.^{20,21,26} This is the case for the system reported by Dufour et al, called "MedihandTrace" (MHT), which meets most of the WHO requirements for an automated hand hygiene systems: continuous recording of HCWs paths and of HH opportunities, addressing the Hawthorne effect (ie, the behavior changes when the subject is observed), implementing real-time reminders, and decreasing time and technical expertise needed.²¹ Huang et al²⁶ used the same system to assess the impact of real-time reminders on HH compliance. They reported an increase of the overall compliance, and they proposed the use of randomized reminders to reduce alarm fatigue of the HCWs. Lastly, Edmisten et al²⁰ implemented a similar technology in 3 community hospitals, and informally observed facility-wide decrease in hospital-acquired infections.

Cleaning and disinfection of hospital environments

Another IPC process reported in the macro areas is the cleaning and disinfection of hospital environments.³² Wang et al reported on

the application of Plan-Do-Check-Act (PDCA) cycle based on artificial intelligence (AI) algorithms in the management of the sterilization of supply rooms. Mainly, they rely on several Long Short-Term Memory (LSTM) neural networks units using 3 types of gating: input gates, forgetting gates, and output gates. LSTMs store and update information through the gating, which works as a fully connected layer. They observed a significant increase in the satisfaction rates and compliance with standardized practices in the group using the PDCA cycle compared to the group using the conventional management method.

Moreover, Khan et al³³ researched the use of robots in hospital settings during COVID-19 pandemic. We focused our attention on the cleaning robots. The paper describes the use and application of different types of cleaning robotic technologies in hospital settings, relying on dry vacuum and mopping to remove germs and pesticides. Other examples of technologies include intelligent navigating vacuum pump, ultra-violet radiation-based device, highly dynamic robotic gripper and sensing system and autonomous heavy-duty cleaning robot. The use of these robots significantly improved the safety and the quality of health care management. Robots were extensively used to control the COVID-19 pandemic, reducing the number of infected patients and casualties.

Infection cluster detection

Two papers examined the use of automated systems for infection cluster detection in the hospital setting.^{28,29} Aghdassi et al²⁹ described an automated cluster alert system (CLAR), based on number of detected isolates, type of pathogen and resistance, sampling material and ward interested. CLAR identified a high number of alerts, validated by IPC physicians. In a similar way, Stachel et al²⁸ implemented an automated surveillance system to detect hospital outbreak, called WHONET-SaTScan, which proved to be a useful addition to their regular IPC program. It works combining a statistical software (WHONET) and a software managing microbiology laboratory data (SaTScan).

Air quality control

Only 1 article dealt with air quality control, focusing their effort in the operating rooms (ORs).³⁰ Colella et al developed a fuzzy inference system (FIS), which assesses the OR air quality and provides real-time alarms, making HCWs aware of potential risk. The risk level is decided by FIS considering 4 parameters, namely particle count, temperature, relative humidity patients and HCWs movements. Typically, FISs are an important part of fuzzy logic systems that perform decision-making and are mainly based on the Mamdani or Sugeno frameworks.³⁴

Correct use of PPE

Another important aspect of IPC is the correct use of PPE, Preda et al³¹ realized an AI-PPE system, with the goal of analyzing donning and doffing with real-time feedback. They validated this technology comparing it to the gold standard, that is, double buddy system. Furthermore, they included in the study participants with heterogeneous visual characteristic (ie, people with different ethnical backgrounds, age, sex, etc.) in order to lower the risk of AI bias.³³

Quality appraisal

The outcome of the MMAT quality analysis can be found in the [Supplementary Table A3](#). Most criteria were met by all studies. Most of the issues rose from the quantitative non-randomized studies. Nine studies did not take into account confounders in the design and analysis, while for 1 study this was unclear. Addressing confounders

is important to avoid misinterpretation of findings, due to spurious associations.^{35,36} Moreover, for 6 studies it was unclear whether the participants were representative of the target population. The lack of clear descriptions of the target population and of the sample can lead to erroneous conclusions, and ultimately to nonresponse bias.³⁷ Potential bias could be present in Akkok et al's study, as they could not prolong its duration due to HCWs' refusal, additionally there is no data available for consultants' HH events.

DISCUSSION

This systematic literature review allowed to highlight 4 points of discussion, arising from the selected paper that led to a wider proposal for the conclusion. Firstly, the main focus of the current literature on the topic is on IPC options, specifically on hand hygiene and UV disinfection rather than the use of robots and IoT. The retrieved and selected articles about robotics are very generic, outlining their general applications, but not giving specific details on the specific field where such solutions may be most required and/or will add most value, benefits and dis-benefit. Overall, HH is the most debated theme and also the one with the most impressive technological advancement (and probably funding). The latest advancement in technologies, AI and IoT could and should be exploited more to allow for autonomous robots and remote operations in health care.

Secondly, HCWs are the first recipients and users of the health technologies reported and described in this systematic literature review. However, it seems that usability engineering principles are often overlooked, as they are not usually involved in the cocreation and codesign of such technologies. Moreover, HCWs lack awareness and education on the importance of complying with the use of such technologies, that if not well prepared could be lacking^{27,38} or that could decrease after the emergencial period.²¹ To accelerate the HCWs response, there might be a need to transform the IPC routine into a habit for them.^{27,39} To monitor and improve HCW's IPC routines in hospitals, Jeanes et al⁴⁰ developed and tested a quality improvement tool, which received good feedback from the participants. When devising and designing such solutions, more attention should be paid towards the fact that HCWs may not like the idea of being tracked and monitored during their already very difficult and stressful professional activities. It is also noteworthy that the intensification of practices stemming from these solutions could have long-term side effects (eg, psychological effects such as mysophobia and compulsive hand washing).⁴¹ Strictly linked to this, but not directly mentioned in the collected literature, there is the need to make doctors more aware and expert of state-of-the-art health care technologies, not only by involving them in the design of the technological solutions used in their professional practice, but also by re-designing their academic education paths. This has already started in some countries (eg, Italy), where medical students will be offered relevant biomedical engineering education spread across their medical studies in order to obtain both a full degree in Medicine and a Bachelor of Science in Biomedical Engineering. The trailblazer for this, in Italy, was Humanitas University in collaboration with Polytechnic University of Milan.⁴² Similarly, other universities are joining forces to offer medical students a path of excellence⁴³ to expand the knowledge of future doctors on new technologies that increasingly impact clinical activity, both diagnostic and therapeutic.

This urgent need to bring doctors closer to advanced health technologies and to improve the robot-human relation also relates to the fast technological enhancing that hospitals are experiencing, on the wake of Health 4.0. In fact, hospitals are undergoing a revolution increasingly becoming more sensorized and robotized, relying on innovative high-tech tools, as portrayed in the selected papers. As mentioned in the introduction, 1 leading example is that of the European (Horizon2020) project ODIN.¹¹ This project is leveraging AI-

based technology to transform the future of health care delivery in leading hospitals in Europe. In particular, its 3 areas of intervention are enhanced hospital workers (ie, exploring how to empower HCWs with appropriate technologies to enhance their skills and support their daily work), enhanced robots (ie, exploring how to automate hospital processes that no longer need humans or can benefit from automation), and enhanced location (ie, exploring how to instrument medical locations for enabling them to proactively support hospital processes). The project relies on 5 European Pilots namely the University medical Centre of Utrecht, the Charité university hospital in Berlin, Medical University of Lodz, University hospital campus biomedico in Rome and Hospital clinico San Carlos in Madrid. There are 7 clinical use cases, that is, aided logistic support, management of medical devices and sites, AI-based support systems for diagnosis, clinical tasks and patient experience, automation of clinical workflows, inpatient remote monitoring, and disaster preparedness.

Furthermore, from the included papers, it is clear that all the emerging technological solutions are not easy to implement, because they are extremely advanced, expensive and their envisioned use environment is up to relevant international standards and minimum requirements. This means that none of them is suitable, as it is, for low-resource settings, that is, contexts that are severely hindered by numerous challenges and characterize both the so-called high-income countries, perhaps in the more rural and peripheral areas, and low-income ones.^{44,45} Although these solutions lack a contextualized and frugal perspective, they should not be overlooked, as they are the main gears pushing forward the frontier of progress, and, as COVID-19 clearly demonstrated, different parts of the worlds can be affected differently and generalist approaches risk being unnecessarily expensive and impossible to achieve uniformly globally.⁴⁶ Nevertheless, it is clear that contextualized and frugal design approaches should be negotiated with the need for uniformity and equality, a utopic idea towards which to strive.

CONCLUSIONS

This systematic literature review demonstrates that IPC practices within hospitals mostly focus on HH and UV devices for disinfection. Not much has been done in regard to the use of IoT, AI, big data technology, robots in the field of IPC within nosocomial settings. This review highlights how most of the literature regarding automation and robots for IPC in hospitals is either outdated or not very impactful, despite the recent COVID-19 pandemic. Nonetheless, the review allowed to highlight 5 main areas that were presented and discussed. Among the main findings, it was noticed that there is no adequate consideration of HCWs in terms of awareness and training with respect to the design and use of health care technologies that impact on their daily work and may have repercussions on their everyday lives. However, their direct involvement in technology co-design and training is strictly necessary, since Health 4.0 is dramatically revolutionizing the way hospitals and HWCs work, leading the digitalization of health care. As mentioned before, one of the trailblazing European projects in this remit is ODIN. Although this is currently mainly concerning high-income countries, which are pushing forward the frontier of progress for finding solutions and novel approaches for future pandemics, research priorities should also be considering how to implement similar or more contextualized options for lower income countries.

FUNDING

Abdulaziz Almuhi is a PhD student sponsored by Prince Sattam bin Abdulaziz University. Leandro Pecchia and Davide Piaggio received support from the University of Warwick with two Warwick

Impact Found grants supported by the EPSRC Impact Accelerator Award (EP/K503848/1 and EP/R511808/1).

SUPPLEMENTARY MATERIALS

Supplementary material associated with this article can be found in the online version at <https://doi.org/10.1016/j.ajic.2023.03.005>.

References

- WHO. *WHO COVID-19 Dashboard*. © World Health Organization; 2020. Accessed October 27, 2022; <https://covid19.who.int/>.
- Maccaro A, Piaggio D, Vignigbè M, Stingl A, Pecchia L. COVID-19 preparedness and social dynamics in a Sub-Saharan Africa country, Benin. *Health Prom Int*. 2022;37:daac105.
- Schwab N, Nienhold R, Henkel M, et al. COVID-19 autopsies reveal underreporting of SARS-CoV-2 infection and scarcity of Co-infections. *Front Med*. 2022;9:868954.
- Lau H, Khosrawipour T, Kocbach P, Ichii H, Bania J, Khosrawipour V. Evaluating the massive underreporting and undertesting of COVID-19 cases in multiple global epicenters. *Pulmonology*. 2021;27:110–115.
- Pecchia L, Piaggio D, Maccaro A, Formisano C, Iadanza E. The inadequacy of regulatory frameworks in time of crisis and in low-resource settings: personal protective equipment and COVID-19. *Health Technol*. 2020;10:1375–1383.
- Müller J, Kretzschmar M. Contact tracing: old models and new challenges. *Infect Dis Model*. 2021;6:222–231.
- Sachs JD, Karim SSA, Akinin L, et al. The Lancet Commission on lessons for the future from the COVID-19 pandemic. *Lancet*. 2022;400:1224–1280.
- WHO. WHO Responds to The Lancet COVID-19 Commission. World Health Organization. Accessed November 3, 2022. <https://www.who.int/news/item/15-09-2022-who-responds-to-the-lancet-covid-19-commission>.
- World Health Organization. Infection prevention and control in the context of coronavirus disease (COVID-19): a living guideline, 25 April 2022: updated chapter: mask use, part 1: health care settings. Accessed April 13, 2023. <https://apps.who.int/iris/handle/10665/353565>.
- Aceto G, Persico V, Pescapé A. Industry 4.0 and health: internet of things, big data, and cloud computing for healthcare 4.0. *J Indust Informat Integr*. 2020;18.
- ODIN. “Odin, Smart Hospitals.” ODIN. Accessed November 3, 2022 <https://odin-smarthospitals.eu/>.
- Yepes-Nuñez J, Urrutia G, Romero-García M, Alonso-Fernández S. The PRISMA 2020 statement: an updated guideline for reporting systematic reviews. *Rev Espan Cardiol (English ed.)*. 2021;74:790–799.
- WHO. Infection Prevention and Control Guidance for Long-Term Care Facilities in the Context of COVID-19: Interim Guidance, 8 January 2021. World Health Organization; 2021.
- Hong QN, Sergi F, Gillian B, Felicity B, et al. The mixed methods appraisal tool (MMAT) version 2018 for information professionals and researchers. *Educ Inform*. 2018;34:285–291.
- Jennie P, Helen R, Amanda S, et al. *Guidance on the conduct of narrative synthesis in systematic reviews. A Product from the ESRC Methods Programme*. 92. Lancaster University; 2006:v1.
- A. Booth, A. Sutton, M. Clowes, and M. Martyn-St James, “Systematic approaches to a successful literature review,” 2021.
- Xu N, Liu C, Feng Y, et al. Influence of the internet of things management system on hand hygiene compliance in an emergency intensive care unit. *J Hosp Infect*. 2021;109:101–106.
- McCalla S, Reilly M, Thomas R, McSpedon-Rai D. An automated hand hygiene compliance system is associated with improved monitoring of hand hygiene. *Am J Infect Control*. 2017;45:492–497.
- McCalla S, Reilly M, Thomas R, McSpedon-Rai D, McMahon LA, Palumbo M. An automated hand hygiene compliance system is associated with decreased rates of health care-associated infections. *Am J Infect Control*. 2018;46:1381–1386.
- Edmisten C, Hall C, Kernizan L, et al. Implementing an electronic hand hygiene monitoring system: lessons learned from community hospitals. *Am J Infect Control*. 2017;45:860–865.
- Dufour J-C, Reynier P, Boudjema S, Aladro AS, Giorgi R, Brouqui P. Evaluation of hand hygiene compliance and associated factors with a radio-frequency-identification-based real-time continuous automated monitoring system. *J Hosp Infect*. 2017;95:344–351.
- Iversen A-M, Kavaliris CP, Hansen R, et al. Clinical experiences with a new system for automated hand hygiene monitoring: A prospective observational study. *Am J Infect Control*. 2020;48:527–533.
- Xu Q, Liu Y, Cepulis D, et al. Implementing an electronic hand hygiene system improved compliance in the intensive care unit. *Am J Infect Control*. 2021;49:1535–1542.
- Xu Q, Liu Y, Cepulis D, et al. Hand hygiene behaviours monitored by an electronic system in the intensive care unit: a prospective observational study. *J Hosp Infect*. 2022;123:126–134.
- Gulsen A, Soysal A, Gul F, et al. Reduction of nosocomial infections in the intensive care unit using an electronic hand hygiene compliance monitoring system. *J Infect Dev Countries*. 2021;15:1923–1928.
- Huang F, Boudjema S, Brouqui P. Three-year hand hygiene monitoring and impact of real-time reminders on compliance. *J Hosp Infect*. 2021;117:111–116.
- Durant DJ, Duvall S, Willis L. Adoption of electronic hand hygiene monitoring systems in NYS hospitals and the impact on hospital-acquired *C. difficile* infection rates. *Am J Infect Contr*. 2020;48:S52.
- Stachel A, Pinto G, Stelling J, et al. Implementation and evaluation of an automated surveillance system to detect hospital outbreak. *Am J Infect Contr*. 2017;45:1372–1377.
- Aghdassi SJS, Britta K, Christin S, et al. Implementation of an automated cluster alert system into the routine work of infection control and hospital epidemiology: experiences from a tertiary care university hospital. *BMC Infect Dis*. 2021;21:1075.
- Colella Y, Valente AS, Rossano L, Trunfio TA, Fiorillo A, Improta G. A fuzzy inference system for the assessment of indoor air quality in an operating room to prevent surgical site infection. *Int J Environ Res Public Health*. 2022;19:3533.
- Preda VA, Jayapadman A, Zacharakis A, et al. Using artificial intelligence for personal protective equipment guidance for healthcare workers in the COVID-19 pandemic and beyond. *Commun Dis Intell*. 2022;46:1–14.
- Wang Y, Zhang S, Chi M, Yu J. A PDCA model for disinfection supply rooms in the context of artificial intelligence to reduce the incidence of adverse events and improve the disinfection compliance rate. *J Healthcare Eng*. 2022;2022:4255751.
- Khan ZH, Siddique A, Lee CW. Robotics utilization for healthcare digitization in global COVID-19 management. *Int J Environ Res Public Health*. 2020;17:3819.
- Mamdani EH, Assilian S. An experiment in linguistic synthesis with a Fuzzy Logic Controller. *Int J Human-Computer Stud*. 1999;51:135–147.
- McNamee R. Confounding and confounders. *Occupat Environ Med*. 2003;60:227.
- VanderWeele TJ, Shpitser I. On the definition of a confounder. *Ann Stat*. 2013;41:196.
- Martínez-Mesa J, González-Chica DA, Duquia RP, Bonamigo RR, Bastos JL. Sampling: how to select participants in my research study? *Anais Brasil Dermatol*. 2016;91:326–330.
- Chowdhury A, Hafeez-Baig A, Gururajan R, Sharif MA. The adoption of mobile technologies in healthcare: the perceptions of healthcare professionals regarding knowledge management practices in developing countries. *International Conference on Human-Computer Interaction*. Springer; 2020:68–76.
- Chamola V, Hassija V, Gupta V, Guizani M. A comprehensive review of the COVID-19 pandemic and the role of IoT, drones, AI, blockchain, and 5G in managing its impact. *IEEE Access*. 2020;8:90225–90265.
- Jeanes A, Coen PG, Drey NS, Gould DJ. Moving beyond hand hygiene monitoring as a marker of infection prevention performance: development of a tailored infection control continuous quality improvement tool. *Am J Infect Control*. 2020;48:68–76.
- Dai H, Milkman KL, Hofmann DA, Staats BR. The impact of time at work and time off from work on rule compliance: the case of hand hygiene in health care. *J Appl Psychol*. 2015;100:846.
- H. University. “6-year Degree Course in Medicine and Biomedical Engineering, Entirely Taught in English, run by Humanitas University in Partnership With Politecnico di Milano.” Accessed November 3, 2022. <https://www.hunimed.eu/course/medtec-school/>.
- U. o. Pavia. “Nasce Meet: Grazie All’incontro Di Quattro Prestigiosi Atenei, Una Formazione Medica All’altezza Delle Nuove Tecnologie.” Accessed November 3, 2022. <http://news.unipv.it/?p=45400>.
- Piaggio D, Castaldo R, Cinelli M, Cinelli S, Maccaro A, Pecchia L. A framework for designing medical devices resilient to low-resource settings. *Global Health*. 2021;17:1–13.
- Di Pietro L, Davide P, Iyabosola O, et al. A framework for assessing healthcare facilities in low-resource settings: field studies in Benin and Uganda. *J Med Biol Eng*. 2020;40:526–534.
- Maccaro A, Piaggio D, Leesurakarn S, et al. On the universality of medical device regulations: the case of Benin. *BMC Health Ser Res*. 2022;22:1–14.